

NATA 2014 FACT SHEET

B BRAUN MEDICAL, INC

ALLENTOWN, PA
HQ: MELSUNGEN, GERMANY

LOCATION

MSA – Allentown-Bethlehem-Easton, PA

Address - 901 Marcon Blvd,
Allentown, PA 18109

TYPE OF FACILITY

Commercial Sterilizer

Major Source

POLLUTANTS

Ethylene Oxide
3.3 tpy (2014)
2.95 tpy (2017)

CANCER RISK

600 in 1 million
(Tract level risk)

2028 in 1 million
(Block level risk)

APPLICABLE RULES

NESHAP – 40 CFR 63 SUBPART O (sterilization)

NSPS - none

SIP - none

COMPLIANCE HISTORY

LAST CAA ON-SITE INSPECTION FROM PADEP 3/14/2018, LAST CAA OFF-SITE INSPECTION 6/12/18. EPA ECHO DATABASE DOES NOT SHOW ANY FORMAL ENFORCEMENT ACTIONS WITHIN THE LAST FIVE YEARS BY EITHER PADEP OR EPA.

PERMITS

8 sterilizers

Controls: scrubber and catalytic oxidation

Title V permit was renewed last on 8/31/2016 and due for renewal on 8/31/2021.

